



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Cal-Corr Services & Repairs
25540 Pennsylvania Road
Taylor, MI 48180

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 02 September 2028

Certificate Number: L2010-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Cal-Corr Services & Repairs

25540 Pennsylvania Road
Taylor, MI 48180
Court Walker 734-942-0900

CALIBRATION

ISO/IEC 17025 Accreditation Granted: **07 July 2026**

Certificate Number: **L2010-1**

Certificate Expiry Date: **02 September 2028**

Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Gas Flow	(0.001 to 849) L/min (0.000 3 to 30) ACFM	0.33 % of reading	Calibration of Air/Gas Flow Meters using Bell Provers
Liquid Flow	(0.001 to 115) L/min (0.000 3 to 30) gpm	0.17 % of reading	Calibration of Liquid Flow Meters using Primary Standard FSC Calibrator
	(0.25 to 190) L/min (0.065 to 50) gpm	0.16 % of reading	Calibration of Liquid Flow Meters using Primary Standard COX 305T Calibrator
	(0.05 to 1 515) L/min (0.013 to 400) gpm	0.16 % of reading	Calibration of Liquid Flow Meters using Primary Standard COX 311AHT Calibrator
Liquid Flow ¹	(6 to 91) L/min (1.7 to 24) gpm	0.24 % of reading	Comparison of Liquid Flow Meters to Coriolis Flow Meter
Calibration of all COX Liquid Flow Weigh and Time Calibrators ¹	(0.05 to 1 515) L/min (0.013 to 400) gpm	0.16 % of reading	Calibration of all Cox Liquid Flow Weigh and Time Calibrators using Primary Standard COX Nozzle Kit

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope



Jason Stine, Vice President

